

0-1" / 0-25mm Electronic Depth / Pit Gauge

Cat # SPI13-749-7



- Large LCD - 0.0005" (0.01mm) Resolution
- ± 0.001 " (0.025mm) Accuracy
- 0.06" Dia Probes
- 2 3/8" Base
- Fitted Case & Battery Included
- Certificate of Calibration Traceable to NIST

Measurement Scale: Inch / Metric

G.A.L. Gage Company



2953 Hinchman Road, Bridgman, Michigan 49106-9501



Your Complete Source for
Testing Equipment Since 1969!

www.BergEng.com
Berg Engineering & Sales Company, Inc.

1-847-577-3980
Info@BergEng.com